

Brussels, XXX [...](2022) XXX draft

ANNEX

ANNEX

to

Commission Delegated Directive

amending, for the purposes of adapting to scientific and technical progress, Annex IV to Directive 2011/65/EU of the European Parliament and of the Council as regards an exemption for mercury in melt pressure transducers for capillary rheometers under certain conditions

EN EN

ANNEX

In Annex IV to Directive 2011/65/EU, the following entry 49 is added:

·49	Mercury in melt pressure transducers for capillary	Applies to category 9
17	rheometers at temperatures over 300°C and	and expires on 31
	pressures over 1000bar	December 2025.'.